

J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN B12 QUANTI MICROLISA
Lot No.	EVB041123
Mfg. Date	2023-11
Expiry. Date	2025-04

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.191	>1.310
S2	100	1.782	
S3	200	1.473	>0.880
S4	400	1.090	
S5	1000	0.629	>0.380
S6	2000	0.382	

Approved / Not Approved

Quality Control


Signature

Date 17-11-2023