

J.MITRA & CO. PVT. LTD.

QC DATA SHEET

VITAMIN B12 QUANTI MICROLISA

Kit	VITAMIN B12 QUANTI MICROLISA
Lot No.	EVB020524
Mfg. Date	2024-05
Expiry. Date	2025-10

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.514	>1.510
S2	100	2.135	
S3	200	1.870	>1.120
S4	400	1.389	
S5	1000	0.762	>0.460
S6	2000	0.454	

✓
Approved / ~~Not Approved~~
Quality Control

Signature *[Handwritten Signature]*

Date 01-06-2024