

J. MITRA & CO. PVT. LTD.

TSH QUANTI MICROLISA

QC DATA SHEET

Kit	TSH QUANTI MICROLISA
Lot No.	ETH030624
Mfg. Date	2024-06
Expiry Date	2026-05

Standard	Concentration (μ IU/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	0.012	
S2	0.5	0.087	
S3	2.5	0.293	>0.180
S4	5.0	0.675	
S5	10.0	1.342	>0.810
S6	20.0	2.186	
S7	40.0	2.795	>1.680

✓
Approved / ~~Not Approved~~

Quality Control


Signature

Date 13-06-2024