

**J.MITRA & CO. PVT. LTD.**

**VITAMIN B12 QUANTI MICROLISA**

**QC DATA SHEET**


Kit	VITAMIN B12 QUANTI MICROLISA
Lot No.	EVB030824
Mfg. Date	2024-08
Expiry. Date	2026-01

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.211	>1.330
S2	100	1.731	
S3	200	1.425	>0.860
S4	400	1.182	
S5	1000	0.634	>0.380
S6	2000	0.365	



Approved / ~~Not Approved~~

Quality Control

  
Signature

Date 08-08-2024