J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB041224
Mfg. Date	2024-12
Expiry. Date	2026-05

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.132	>1.280
S2	100	1.904	
S3	200	1.521	>0.910
S4	400	1.138	
S5	1000	0.605	>0.360
S6	2000	0.361	

Approved / ~~Not Approved~~

Quality Control

Signature

Date 25-12-2024