J. MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN D QUANTI MICROLISA**
Lot No.	EVD031025
Mfg. Date	2025-10
Expiry Date	2027-03

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.299	>1.380
S2	4	1.918	
S3	10	1.523	>0.910
S4	25	1.098	
S5	60	0.419	
S6	130	0.075	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration (ng/ml)
Low	20	15-25	18.15
High	50	40-60	52.83

Approved / ~~Not Approved~~

Quality Control

Signature

Date: 31-10-2025