J.MITRA & CO. PVT. LTD.

hs-CRP QUANTI MICROLISA

QC DATA SHEET

Kit	**hs-CRP Quanti Microlisa**
Lot No.	ECP020721
Mfg. Date	2021-07
Expiry. Date	2022-12

Standard	Concentration (mg/L)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	0.026	
S2	0.5	0.093	>0.06
S3	2	0.240	
S4	5	0.671	>0.4
S5	15	2.187	
S6	30	2.651	>1.59

QC **Approved** / Not Approved

Quality Control

Signature

Date 17/07/2021