J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA


DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB020521
Mfg. Date	2021-05
Expiry. Date	2022-10

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range**
S1	0	1.992	>1.2
S2	100	1.746	
S3	200	1.444	>0.87
S4	400	1.045	
S5	1000	0.495	>0.3
S6	2000	0.311	

Approved / ~~Not Approved~~

Quality Control



Signature

Date 02/06/2021