J J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

DATA SHEET

Kit	VITAMIN D QUANTI MICROLISA
Lot No.	EVD020421
Mfg. Date	2021-04
Expiry. Date	2022-06

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.662	>1.46
S2	4	2.155	>1.19
S3	10	1.972	>1.08
S4	25	1.246	>0.69
S5	60	0.771	>0.42
S6	130	0.272	>0.15

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	20	15-25	18.29
High	50	40-60	58.5

Approved / Not Approved

Signature

Date 30/04/2021