J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB010321
Mfg. Date	2021-03
Expiry. Date	2022-08

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range**
S1	0	2.122	>1.49
S2	100	1.751	
S3	200	1.520	>1.06
S4	400	1.109	
S5	1000	0.499	>0.35
S6	2000	0.294	

Approved / Not Approved

Quality Control

Signature

Date 08/03/2021