J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB030721
Mfg. Date	2021-07
Expiry. Date	2022-12

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range**
S1	0	2.294	>1.38
S2	100	1.834	
S3	200	1.548	>0.93
S4	400	1.040	
S5	1000	0.653	>0.39
S6	2000	0.374	

Approved / ~~Not Approved~~

Quality Control

Signature

Date 26/07/2021