J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB041021
Mfg. Date	2021-10
Expiry. Date	2023-03

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range**
S1	0	2.325	>1.400
S2	100	1.787	
S3	200	1.573	>0.940
S4	400	0.906	
S5	1000	0.469	>0.280
S6	2000	0.291	

Approved / Not Approved

Quality Control

Signature

Date 29/10/2021