J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN D QUANTI MICROLISA**
Lot No.	EVD041021
Mfg. Date	2021-10
Expiry. Date	2022-12

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.433	>1.460
S2	4	1.995	
S3	10	1.550	>0.930
S4	25	1.123	
S5	60	0.650	>0.390
S6	130	0.138	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	18.87	15-25	19.52
High	47.03	40-60	51.22

Approved / Not Approved

Signature

Date 18/10/2021