J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB010222
Mfg. Date	2022-02
Expiry. Date	2023-07

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range of O.D.**
S1	0	2.346	>1.410
S2	100	1.794	
S3	200	1.532	>0.920
S4	400	1.006	
S5	1000	0.514	>0.310
S6	2000	0.315	

Approved / ~~Not Approved~~

Quality Control

Signature

Date 05/02/2022