J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB020422
Mfg. Date	2022-04
Expiry. Date	2023-09

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range of O.D.**
S1	0	2.304	>1.380
S2	100	2.005	
S3	200	1.665	>1.000
S4	400	1.190	
S5	1000	0.536	>0.320
S6	2000	0.332	

Approved / Not Approved

Quality Control

Signature

Date 14/04/2022