J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN D QUANTI MICROLISA**
Lot No.	EVD020422
Mfg. Date	2022-04
Expiry. Date	2023-09

Standard	**Concentration (ng/ml)**	**Avg. O.D. at 450/630 nm**	**Acceptable range**
S1	0	2.295	>1.380
S2	4	1.778	
S3	10	1.564	>0.940
S4	25	0.849	
S5	60	0.663	
S6	130	0.131	

Standard	**Concentration (ng/ml)**	**Acceptable range (ng/ml)**	**Observed Concentration(ng/ml)**
Low	20	15-25	22.28
High	50	40-60	51.87

Approved / Not Approved

Signature

Date 20/04/2022