J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN D QUANTI MICROLISA
Lot No.	EVD030522
Mfg. Date	2022-05
Expiry. Date	2023-10

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.582	>1.550
S2	4	2.065	
S3	10	1.581	>0.950
S4	25	1.008	
S5	60	0.555	
S6	130	0.115	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	20	15-25	18.37
High	50	40-60	49.02

Approved / Not Approved

Signature

Date 17/05/2022