J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN D QUANTI MICROLISA**
Lot No.	EVD040822
Mfg. Date	2022-08
Expiry. Date	2024-01

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.394	>1.440
S2	4	2.006	
S3	10	1.630	>0.980
S4	25	1.117	
S5	60	0.544	
S6	130	0.088	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	20	15-25	22.73
High	50	40-60	52.60

Approved / Not Approved

Signature

Date 13/08/2022