J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB050822
Mfg. Date	2022-08
Expiry. Date	2024-01

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.353	>1.410
S2	100	1.717	
S3	200	1.665	>1.000
S4	400	1.148	
S5	1000	0.624	>0.370
S6	2000	0.379	

Approved / Not Approved

Quality Control

Signature

Date 02/09/2022