J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB060922
Mfg. Date	2022-09
Expiry. Date	2024-02

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.414	>1.450
S2	100	2.090	
S3	200	1.553	>0.930
S4	400	1.156	
S5	1000	0.680	>0.410
S6	2000	0.358	

Approved / Not Approved

Quality Control

Signature

Date 28/09/2022