J.MITRA & CO. PVT. LTD.

hs-CRP QUANTI MICROLISA

QC DATA SHEET

Kit	**hs-CRP Quanti Microlisa**
Lot No.	ECP021122
Mfg. Date	2022-11
Expiry. Date	2024-04

Standard	Concentration (mg/L)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	0.017	
S2	0.5	0.072	>0.04
S3	2	0.220	
S4	5	0.680	>0.4
S5	15	2.178	
S6	30	2.670	>1.6

QC **Approved** / Not Approved

Quality Control

Signature

Date 30/11/2022