J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB020423
Mfg. Date	2023-04
Expiry. Date	2024-09

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.272	>1.360
S2	100	1.828	
S3	200	1.657	>0.990
S4	400	1.188	
S5	1000	0.663	>0.400
S6	2000	0.347	

Approved / Not Approved

Quality Control

Signature

Date 20/04/2023