J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	**VITAMIN B12 QUANTI MICROLISA**
Lot No.	EVB030823
Mfg. Date	2023-08
Expiry. Date	2025-01

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.231	>1.340
S2	100	1.874	
S3	200	1.629	>0.980
S4	400	1.208	
S5	1000	0.511	>0.310
S6	2000	0.325	

Approved / Not Approved

Quality Control

Signature

Date 29/08/2023