J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN D QUANTI MICROLISA
Lot No.	EVD041223
Mfg. Date	2023-12
Expiry. Date	2025-05

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.331	>1.400
S2	4	1.964	
S3	10	1.640	>0.980
S4	25	1.221	
S5	60	0.586	
S6	130	0.125	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	20	15-25	22.73
High	50	40-60	55.10

Approved / Not Approved

Quality control

Signature

Date 30-12-2023