

MALARIA Ag MICROLISA

Microwell ELISA Test for detecting infection with plasmodium species
(*P. falciparum*/*P. vivax*/*P. malariae*/*P. ovale*) based on pLDH Antigen in Human Blood

1. SUMMARY AND EXPLANATION OF THE TEST

Malaria is a serious, sometimes fatal, parasitic disease characterized by fever, chills, and anaemia and is caused by a parasite that is transmitted from one human to another by the bite of infected Anopheles mosquitoes. There are four kinds of malaria parasite that can infect human: *Plasmodium falciparum*, *P. vivax*, *P. ovale* and *P. malariae*. The disease now occurs in more than 90 countries worldwide, and it is estimated that there are over 500 million clinical cases and 2.7 million malaria-caused deaths per year. At present, malaria is diagnosed microscopically by looking for the parasites in a drop of blood. Malaria Ag Microlisa is a microwell ELISA test for the diagnosis of infection with all four Plasmodium Species (*P. falciparum*/*P. vivax*/*P. malariae*/*P. ovale*) in human whole blood. Malaria Ag Microlisa is specially designed to exclude infected blood to prevent transfusion acquired malaria and is used to monitor of anti-malarial therapy.

2. INTENDED USE














Malaria Ag Microlisa is an in-vitro qualitative enzyme immunoassay for detecting infection with plasmodium species (*P. falciparum*/*P. vivax*/*P. malariae*/*P. ovale*) based on pLDH antigen in human blood. It is intended for screening of blood donors or other individuals at risk for malaria infection and for clinical diagnostic testing.

3. PRINCIPLE OF THE TEST

Malaria Ag Microlisa test is an enzyme immunoassay based on sandwich ELISA. Anti-pLDH antibodies have been coated onto the wells of the microwell plate provided. Blood sample including controls are pipetted into these wells. During the first incubation, the plasmodium species pLDH antigen from samples and coated antibody react. After washing the enzyme conjugate (anti-pLDH antibody HRPO) is added and incubated. Washing is done again to remove all the unbound enzyme conjugate. A substrate solution which acts on the bound enzyme is added to induce a coloured reaction product. The intensity of this coloured product is directly proportional to the concentration of pLDH present in the samples.

4. DESCRIPTION OF SYMBOLS USED

The following are graphical symbols used in or found on J. Mitra diagnostic products and packing. These symbols are the most common ones appearing on medical devices and their packing. They are explained in more detail in the European Standard EN ISO 15223-1:2016.

 Manufactured By	 <i>In vitro</i> diagnostic medical device
 No. of tests	 See Instruction for use
 Lot Number Batch Number	 Temperature Limitation
 Manufacturing Date	 Caution See instruction for use
 Expiry Date	 Catalogue Number
 Do not use if package is damaged	 Keep away from sunlight
 Keep Dry	



5. PACK SIZE

- 96 Tests

6. COMPONENTS IN EACH MALARIA Ag MICROLISA KIT

Store all components at 2-8°C when not in use. Expiry date on the kit indicates that beyond which the kit should not be used.

Malaria Ag Microlisa Strip Plates	12 Strips (12 X 8 wells) Breakway microwells coated with anti-pLDH antibody packed in a pouch with desiccant.
Sample Diluent	1 Bottle (15 ml.) Buffer containing protein stabilizers and antimicrobial agents as preservative.
Enzyme Conjugate Concentrate (51x)	1 Vial (0.3 ml.) Anti-pLDH antibody conjugated with horseradish peroxidase with protein stabilizers and preservatives.

Conjugate Diluent	1 Bottle (15 ml.) Buffer containing protein stabilizers and preservatives.
Wash Buffer Concentrate (25x)	1 Bottle (50 ml.) PBS with surfactant. Dilute 1:25 with distilled water before use.
TMB Substrate	1 Bottle (10 ml.) To be diluted with TMB diluent before use.
TMB Diluent	1 Bottle (10 ml.) Buffer solution containing H ₂ O ₂ with preservative
Control - 	1 Vial (2.0 ml.) Ready to use, normal bovine serum albumin negative for Malaria, contains sodium azide as preservative.
Control + 	1 Vial (1.0 ml.) Ready to use, pLDH antigen positive for Malaria, contains sodium azide as preservative.
Stop Solution	1 Bottle (15 ml.) Ready to use, 1N sulfuric acid.
Plate Sealers	Adhesive backed sheets for sealing microwell plate/strips.


7. ADDITIONAL MATERIAL AND INSTRUMENTS REQUIRED

- Micropipettes and microtips.
- Elisa reader
- Distilled or deionized water
- Graduated Cylinders, for reagent dilution
- Disposable gloves
- Glassware
- Timer
- Elisa washer
- Incubator 37°C
- Vortex Mixer
- Paper towels or absorbent tissue

8. SPECIMEN COLLECTION & PREPARATION

1. Collect the whole blood in a clean container (containing EDTA, citrate or heparin) by venipuncture. Fresh anticoagulated whole blood samples are preferred for testing as they perform best when tested immediately after collection. If samples are not immediately tested, they should be stored at 2-8°C for not more than 3 days or frozen at minus (-20°C) for long term storage, otherwise false / erroneous results may be obtained.
2. Repeated freeze/ thaw of the specimen should be avoided as it will lead to erroneous results.
3. Specimens should be free of microbial contamination.

10. WARNING & PRECAUTION

 **CAUTION:** THIS KIT CONTAINS MATERIALS OF HUMAN ORIGIN. NO TEST METHOD CAN OFFER COMPLETE ASSURANCE THAT HUMAN BLOOD PRODUCTS WILL NOT TRANSMIT INFECTION. NEGATIVE CONTROL, POSITIVE CONTROL & ALL THE SAMPLES TO BE TESTED SHOULD BE HANDLED AS THOUGH CAPABLE OF TRANSMITTING INFECTION.

1. The use of disposable gloves and proper biohazardous clothing is STRONGLY RECOMMENDED while running the test.
2. In case there is a cut or wound in hand, DO NOT PERFORM THE TEST.
3. Do not smoke, drink or eat in areas where specimens or kit reagents are being handled.
4. Tests are for *in vitro* diagnostic use only and should be run by competent person only.
5. Do not pipette by mouth.
6. All materials used in the assay and samples should be decontaminated in 5% sodium hypochlorite solution for 30-60 min. before disposal or by autoclaving at 121°C at 15psi for 60 min. Do not autoclave materials or solution containing sodium hypochlorite. They should be disposed off in accordance with established safety procedures.
7. Wash hands thoroughly with soap or any suitable detergent, after the use of the kit. Consult a physician immediately in case of accident or contact with eyes, in the event that contaminated material are ingested or come in contact with skin puncture or wounds.
8. Spills should be decontaminated promptly with Sodium Hypochlorite or any other suitable disinfectant.

9. Sample Diluent and controls contains Sodium Azide as a preservative. If these material are to be disposed off through a sink or other common plumbing systems, flush with generous amounts of water to prevent accumulation of potentially explosive compounds. In addition, consult the manual guideline "Safety Management No. CDC-22", Decontamination of Laboratory Sink Drains to remove Azide salts" (Centre for Disease Control, Atlanta, Georgia, April 30, 1976).
10. Follow standard biosafety guidelines for handling & disposal of potentially infective material.
11. ELISA Reader & micropipettes used in testing should be calibrated at regular interval to ensure accurate results.

11. PRECAUTIONS FOR USE

Optimal assay performance requires strict adherence to the assay procedure described in the manual.

1. Do not use kit components beyond the expiration date which is printed on the kit.
2. Bring all the reagents & samples to room temperature (20-30°C) before use.
3. Do not combine reagents from different batches, as they are optimised for individual batch to give best results.
4. Avoid microbial contamination of reagents. The use of sterile disposable tips is recommended while removing aliquots from reagent bottles.
5. Due to interchange of caps the reagents may get contaminated. Care should be taken while handling the reagents to avoid contamination of any sort.
6. Use freshly collected, clean blood samples for assay.
7. Use a separate tip for each sample and then discard it as biohazardous waste.
8. All pipetting steps should be performed with utmost care and accuracy. Cross contamination between reagents and samples will invalidate results.
9. Do not allow microwells to dry once the assay has started.
10. Run negative and positive controls in each assay to evaluate validity of the kit.
11. Incubation time should not vary by more than ± 2 min.
12. Prevent evaporation during sample incubation by covering the strips with strip sealer. Remove sealer before washing.
13. Distilled or deionised water must be used for wash buffer preparation.
14. Thorough washing of the wells is critical to the performance of the assay. Overflowing of reagents or washing to adjacent wells must be prevented during washing, which may lead to incorrect results due to carry over effect.
15. Take care while preparing working substrate solution as Bottle of TMB Substrate & Diluent are of same size.
16. Prepare working substrate solution just 10 minutes prior to adding in the wells.
17. If blue colour or white particles appear in working substrate solution then do not use it. Take fresh containers and tips and prepare it again.
18. Use separate tips for TMB Substrate and TMB diluent.
19. Avoid strong light exposure during the assay.
20. Ensure that the microwell strips are levelled in the strip holder. Before reading, wipe the bottom of the microwell strips carefully with soft, absorbent tissue to remove any moisture.
21. If available, a microwell reader which contains a reference filter with settings at 620 or 630 nm should be used. Use of a reference filter minimises interference due to microwells that are opaque, scratched or irregular. However, if a reference filter is unavailable, the absorbance may be read at 450 nm without a reference filter.
22. In case of any doubt the run should be repeated.

12. PREPARATION OF REAGENTS

Prepare the following reagents before or during assay procedures. Reagents and samples should be at room temperature (20-30°C) before beginning the assay and can remain at room temperature during testing. Return reagents to 2-8°C after use. All containers used for preparation of reagents must be cleaned thoroughly and rinsed with distilled or deionized water. Prewarm the incubator at 37°C.

1. Malaria Ag Microlisa Strip:

Bring plate pouch to room temperature (20-30°C) before opening to prevent condensation on the microwell strips.

- a. Break-off the required number of strips needed for the assay and place in the well holder. Take the strip holder with the required number of strips, taking into account that three negative & two positive controls should be included in the run while opening the fresh kit. However for one or two strips, two negative and one positive control and for more strips at least three negative and two positive controls should be included in each subsequent runs.

- b. **Unused wells should be stored at 2-8°C, with dessicant in a aluminium pouch with clamp & rod. Microwells are stable for 30 days at 2-8°C from the date of opening of sealed pouch, when stored with desicant along with clamp & rod. Caution:** Handle microwell strip with care. Do not touch the bottom exterior surface of the wells.

2. Preparation of Wash Buffer:

- a) Check the buffer concentrate for the presence of salt crystals. If crystals are present in the solution, resolubilize by warming at 37°C until all crystals dissolve.
- b) Prepare at least 50 ml. (2ml. concentrated buffer with 48 ml. water) of buffer for each microlisa strip used. Mix well before use.
- c) **Mix 20 ml. 25X wash buffer concentrate with 480 ml. of distilled or deionized water. Wash buffer is stable for 2 months when stored at 2-8°C.**

3. Preparation of Working Conjugate:

Dilute Enzyme conjugate concentrate 1:51 in conjugate diluent. **Do not store working conjugate.** Prepare a fresh dilution for each assay in a clean glass vessel. Determine the quantity of working conjugate solution to be prepared from the table below. Mix solution thoroughly before use.

No. of Strips	1	2	3	4	5	6	7	8	9	10	11	12
No. of Wells	8	16	24	32	40	48	56	64	72	80	88	96
Enzyme Conjugate Concentrate (μ l)	20	40	60	80	100	120	140	160	180	200	220	240
Conjugate Diluent (ml)	1.0	2.0	3.0	4.0	5.0	6.0	7.0	8.0	9.0	10.0	11.0	12.0

Note: In case any precipitate is found in conjugate diluent, it should be allowed to settle and the supernatant can be used for the test. The precipitate does not interfere with the working of the kit.

4. Preparation of working substrate solution :

Mix TMB substrate and TMB Diluent in 1:1 ratio to prepare working substrate.

No. of Strips	1	2	3	4	5	6	7	8	9	10	11	12
No. of Wells	8	16	24	32	40	48	56	64	72	80	88	96
TMB Substrate (ml)	0.5	1.0	1.5	2.0	2.5	3.0	3.5	4.0	4.5	5.0	5.5	6.0
TMB Diluent (ml)	0.5	1.0	1.5	2.0	2.5	3.0	3.5	4.0	4.5	5.0	5.5	6.0

Do not store working substrate. Prepare a fresh dilution for each assay in a clean plastic/glass vessel. Determine the quantity of working substrate solution to be prepared from table. Mix solution thoroughly before use. Discard unused solution. A deep blue color present in the substrate solution indicates that the solution has been contaminated and must be discarded.

13. WASH PROCEDURE:

1. Incomplete washing will adversely affect the test outcome.
2. Aspirate the well contents completely into a waste container. Then fill the wells completely with wash buffer avoiding overflow of buffer from one well to another and allow to soak (approx. 30 seconds). Aspirate completely and repeat the wash and soak procedure 5 additional times for a total of 6 washes.
3. Automated washer if used should be well adjusted to fill each well completely without over filling or under filling.
4. Tap upside down on absorbent sheet till no droplets appear on the sheet, taking care not to dislodge the wells.

14. TEST PROCEDURE

Once the assay has started, complete the procedure without interruption. All the reagents should be dispensed in the centre of the well and the tip of the pipette should not touch the wall of the microwell.

Fit the strip holder with the required number of Malaria Ag Microlisa strips. The sequence of the procedure must be carefully followed. From well A-1 arrange all controls in a horizontal or vertical configuration. Configuration is dependent upon reader software.

1. Add 100 μ l sample diluent in each well.
2. Add 25 μ l Negative Control in each well no. A-1, B-1 & C-1 respectively.
3. Add 25 μ l Positive Control in D-1 & E-1 wells.
4. Add 25 μ l blood sample in each well starting from F-1.
5. Apply cover seal.
6. Incubate at 37°C \pm 2°C for 60 min. \pm 2 min.

- While the samples are incubating, prepare Working Wash Solution and Working Conjugate as specified in Preparation of Reagents.
- Take out the plate from the incubator after the incubation time is over and, wash the wells 6 times with Working Wash solution according to the wash procedure given in the previous section (wash procedure).
- Add 100 μ l of Working Conjugate Solution in each well.
- Apply cover seal.
- Incubate at 37°C \pm 2°C for 30 min. \pm 2 min.
- Aspirate and wash as described in step no. 8.
- Add 100 μ l of working substrate solution in each well.
- Incubate at room temperature (20 - 30°C) for 30 min. in dark.
- Add 100 μ l of stop solution.
- Read absorbance at 450 nm at 30 minutes in ELISA READER. (Bichromatic absorbance measurement with a reference wavelength 600 - 650 nm is recommended when available).

15. SUMMARY OF PROCEDURE

Add Sample Diluent		100 μ l
Add control & Blood samples		25 μ l
Cover the plate & incubate		60 mins. at 37°C
Wash		6 Cycles
Prepare working conjugate		No of Strips Enz. conc. (20 40 60 80 100 120 140 160 180 200 220 240 μ l) Diluent (ml.) (1 2 3 4 5 6 7 8 9 10 11 12)
Add Conjugate		100 μ l
Cover the plate & incubate		30 mins. at 37°C
Wash		6 Cycles
Prepare working TMB Substrate		No of Strips TMB (0.5 1.0 1.5 2.0 2.5 3.0 3.5 4.0 4.5 5.0 5.5 6.0) Substrate (ml) (0.5 1.0 1.5 2.0 2.5 3.0 3.5 4.0 4.5 5.0 5.5 6.0) TMB Diluent (ml.)
Add Substrate		100 μ l
Incubate in dark		30 mins. at Room Temp.
Add Stop Solution		100 μ l
Read Results		450 nm./630 nm.

16. CALCULATION OF RESULTS

Abbreviations

- NC - Absorbance of the Negative Control
 NC \bar{x} - Mean Negative Control
 PC - Absorbance of the Positive Control
 PC \bar{x} - Mean Positive Control

TEST VALIDITY:

Negative Control Acceptance Criteria:

NC must be \leq 0.150. If it is not so, the run is invalid and must be repeated.

Positive Control Acceptance Criteria:

- PC must be \geq 1.00

CUT OFF VALUE

Cut-off value can be determined by using the following formula:

$$\text{Cut-off Value} = \text{NC}\bar{x} + 0.100$$

Where NC \bar{x} is mean absorbance (O.D) of Negative Control.

e.g. 0.041 + 0.100 = 0.141

17. INTERPRETATION OF RESULTS

- Test specimens with absorbance value less than the cut off value are non-reactive and may be considered as negative for plasmodium species pLDH antigen and interpret results as malaria negative.
- Test specimens with absorbance value greater than or equal to the cut off value are reactive for pLDH antigen by Malaria Ag Microlisa and interpret results as malaria positive.

Note: Test specimens with absorbance value within 10% below the cut off should be considered suspect for the presence of antigen, should be retested in duplicate.

- Specimens with absorbance value equal to or greater than the cut off value are considered initially reactive by the criteria of Malaria Ag Microlisa. Original specimen should be retested in duplicate.
- If both duplicate retest sample absorbance value is less than cut off value, the specimen is considered non reactive.
- If any one of the-duplicates retest sample absorbance value is equal to or greater than the cut off or both duplicate retest value are equal to or greater than the cut off, the specimen is considered reactive by the criteria of Malaria Ag Microlisa. Further confirmation by other EIA assays is recommended.
- Specimens which are not repeatedly reactive, may have shown colour due to:
 - Carry over of a highly reactive sample due to contamination of pipette tips.
 - Substrate contamination
 - Inadequate wash or aspiration during wash procedure.

18. LIMITATIONS OF THE ASSAY

1. Malaria Ag Microlisa test kit is designed for testing plasmodium species pLDH antigen in human blood. Other body fluids and pooled samples are not recommended in this assay. Any result derived from the test of any other body fluid or from test of pooled serum/plasma may not be interpreted correctly based on the current criteria. **This is only a screening test.** All samples detected reactive must be confirmed by using other EIA test, microscopy or PCR. Therefore for a definitive diagnosis, the patient's clinical history, symptomatology as well as serological data should be considered. The results should be reported only after complying with the above procedure.

2. **This is only a screening test.** The results of the test are to be interpreted within the epidemiological, clinical and therapeutic context. When it seems indicated, the parasitological techniques of reference should be considered (microscopic examination of the thick smear and thin blood films).

3. The performance of Malaria Ag Microlisa has not been evaluated with P. malarie and P. ovale positive samples.

19. PERFORMANCE CHARACTERISTICS

The **Malaria Ag Microlisa** test kit has been evaluated in-house with malaria positive and negative clinical whole blood samples and compared with microscopic examination. The evaluation also included cross-reacting samples; Dengue, Rheumatoid factor, Leptospiira, HIV, HCV, HBV, M. tuberculosis, Syphilis, Brucella, Scrub typhus positive samples. The results obtained are as follows:

Sample	Total no. of samples tested	Malaria Ag Microlisa		Sensitivity (%)	Specificity (%)
		Positive	Negative		
Malaria Negative	100	0	100	-	100
P. falciparum Positive	10	10	0	100	-
P. vivax Positive	30	30	0	100	-

Analytical Sensitivity: The Malaria Ag Microlisa test can detect parasitemia level of \geq 20 parasites per μ l of blood for both P. falciparum (pLDH) and P. vivax (pLDH).

Analytical Specificity: There was no significant interference in the test results of Malaria Ag Microlisa when potentially interfering samples; high bilirubin (38 mg/dl), high Hb (18 mg/dl), High triglycerides (975 mg/dl), high albumin (20 mg/dl) were added to the test specimen with much higher level than in normal blood.

Precision :

Intra-Assay: Repeatability has been determined by testing 10 replicates of two weak (1 each of Pf & Pv) positive and two strong positive (1 each of Pf & Pv) and 2 negative samples run on the same lot. The C.V (%) for all samples is < 10%.

Inter-Assay: Reproducibility has been determined by testing 10 replicates of two weak (1 each of Pf & Pv) positive and two strong positive (1 each of Pf & Pv) and 2 negative samples in 10 different run on different day. The C.V (%) for all samples is < 15%.

20. LIMITED EXPRESSED WARRANTY DISCLAIMER

The manufacturer limits the warranty to the test kit, as much as that the test kit will function as an in-vitro diagnostic assay within the limitations and specifications as described in the product instruction-manual, when used strictly in accordance with the instructions contained therein. The manufacturer disclaims any warranty expressed or implied including such expressed or implied warranty with respect to merchantability, fitness for use or implied utility for any purpose. The manufacturer's liability is limited to either replacement of the product or refund of the purchase price of the product and in no case liable to for claim of any kind for an amount greater than the purchase price of the goods in respect of which damages are likely to be claimed.

The manufacturer shall not be liable to the purchaser or third parties for any injury, damage or economic loss, howsoever caused by the product in the use or in the application there of.

21. REFERENCES

1. Alfons Jimenez[†], Roxanne R. Rees-Channer[†], Rushini Perera, Dionicia Gamboa, Peter L. Chiodini, Xavier C. Ding, Malaria Journal 201716:128
2. Jessica Maltha Email author, Issa Guiraud, Palpouguini Lompo, Bérenger Kaboré, Philippe Gillet, Jan Jacobs, Malaria Journal 201413:20
3. Jin Woo Jang[†], Chi Hyun Cho, Seung Lim Email author, Malaria Journal 201312:181
4. Pascal S Atchade[†], Cécile Doderer-Lang[†], Nicodème Chabi, Sylvie Perrotey, Ermanno Candolfi[†] Email author, Malaria Journal 201312:279

22. TROUBLE SHOOTING CHART

PROBLEM	POSSIBLE CAUSE	SOLUTION
1. No colour developed at the end of assay	a) Any one reagent has been added in wrong sequence.	Follow the procedure meticulously & repeat assay.
	b) Inactivated conjugate, improper storage	Check storage of enzyme conjugate and it should be free of any contamination.
	c) Microplate inactivated, due to improper storage	Keep unused strips in aluminium poly pouch with the dessicant pouch inside and properly closed with clamp & rod.
	d) Inactivated substrate, improper storage or preparation	Use freshly prepared substrate solution. Recheck procedure, repeat assay
	e) Omission of any step in test procedure	Follow the procedure meticulously & repeat assay.
	f) Incorrect (low) incubator temperature, timing or pipetting	Check incubator temperature, procedure & repeat assay
	g) Improper preparation of enzyme conjugate (dilution error) improper mixing of reagents.	Check procedure & repeat assay
2. High O.D. value of Negative control	a) Plate not stopped after 30 minutes of adding stop solution	Follow the procedure meticulously & repeat assay.
	b) Same microtip used for Positive and negative controls	Change micropipette tips while addition of negative/ positive control
	c) Nonspecific attachment/ binding of other reagent	If plates get scratches/ aberrations during washing, non specific proteins may bind while addition of next step.
3. Too much colour in all wells of the plate (high background)	a) Contaminated substrate	Check substrate (TMB Diluent) it should be colourless. If blue in colour then discard and use clean disposable container.
	b) Contaminated washing solution (1X). Poor quality of water used for diluting wash buffer conc.	Check the container and quality of water used for dilution. Use of distilled water is preferred.
	c) Over incubation of substrate and delay in addition of stop solution.	Follow the procedure meticulously.

PROBLEM	POSSIBLE CAUSE	SOLUTION
4. Poor reproducibility	d) Insufficient washing. i) Washing not consistent ii) Filling volume not sufficient. iii) Insufficient no. of wash cycles. iv) Contaminated wash device	Check wash device, fill the well close to the top. After washing, blot the microwells on absorbent tissue. Follow wash protocol meticulously
	e) Use of wash solution from other manufacturer.	Use only Malaria Ag MicroIisa wash solution.
	f) Working substrate not protected from light	Incubate the plate in dark after addition of substrate.
	a) Washing problems.	Check all 8 ports/ manifold for uniform flow of wash buffer. If there are blockage, clen the ports.
	b) Uncalibrated pipettes or tips not well fitted, improper pipetting/ dispensing.	Use only calibrated pipettes with well fitted tips & pipette carefully without bubbling.
	c) Interference in optical pathway due to Air bubbles.	Clean or dry the bottom of microwells, check for bubbles and repeat the readings.
5. False Positive	Beside 3a, b, c, d, e & f incorrect interpretation and calculation of final results	Check the calculation part given in the insert and correctly interpret.
6. False Negative/ low O.D. of Positive control & positive sample	a) Inadequate addition of substrate/conjugate solution.	Follow the procedure meticulously & repeat assay.
	b) Kit expired, reagent of different kit used.	Check the expiry of the kit before use.
	c) White particles in working substrate solution.	Discard the substrate and prepare the working substrate again in fresh tube.
	d) Uncalibrated pipettes, improper pipetting.	Use only calibrated pipettes with well fitted tips & pipette carefully without bubbling.
	e) Deterioration of Enzyme conjugate	Check storage of Enzyme conjugate. It shall be stored at 2-8°C.
	f) Stop solution is added before 30 minutes. Reaction terminated before 30 minutes.	Follow the test procedure meticulously.
	g) O.D. taken at incorrect wavelength.	Read O.D. values at 450 nm and 630 nm.
	h) Incorrect (low) incubator temperature, timing or pipetting	Check incubator temperature, procedure & repeat assay

in vitro diagnostic Reagent, not for medicinal use



J. Mitra & Co. Pvt. Ltd.

A-180-181, Okhla Ind. Area, Ph.-1, New Delhi-110 020, INDIA
Ph.: +91-11-47130300, 47130500
e-mail: jmitra@jmitra.co.in Internet: www.jmitra.co.in

R-02

VER-01

MINE/MA/006
Rev. Date: Sep.-21