

J.MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN D QUANTI MICROLISA
Lot No.	EVD010225
Mfg. Date	2025-02
Expiry. Date	2026-07

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.419	>1.450
S2	4	1.929	
S3	10	1.615	>0.970
S4	25	1.255	
S5	60	0.507	
S6	130	0.058	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration(ng/ml)
Low	20	15-25	23.29
High	50	40-60	55.93

✓
Approved / ~~Not Approved~~

Quality control


Signature

Date 01-03-2025