

J. MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN B12 QUANTI MICROLISA
Lot No.	EVB020725
Mfg. Date	2025-07
Expiry Date	2026-12

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.577	>1.550
S2	100	1.785	
S3	200	1.529	>0.920
S4	400	1.265	
S5	1000	0.541	>0.320
S6	2000	0.317	

✓
Approved / ~~Not Approved~~

Quality Control

Signature 

Date: 28-07-2025