

J. MITRA & CO. PVT. LTD.

VITAMIN D QUANTI MICROLISA

QC DATA SHEET

Kit	VITAMIN D QUANTI MICROLISA
Lot No.	EVD020926
Mfg. Date	2026-09
Expiry Date	2028-02

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range
S1	0	2.224	>1.330
S2	4	1.866	
S3	10	1.621	>0.970
S4	25	1.237	
S5	60	0.489	
S6	130	0.075	

Standard	Concentration (ng/ml)	Acceptable range (ng/ml)	Observed Concentration (ng/ml)
Low	20	15-25	19.84
High	50	40-60	50.50

✓
Approved / ~~Not Approved~~

Quality Control


Signature

Date:- 24-09-2026