

**J.MITRA & CO. PVT. LTD.**

**T3 QUANTI MICROLISA**

**QC DATA SHEET**

Kit	T3 QUANTI MICROLISA
Lot No.	ETA031123
Mfg. Date	2023-11
Expiry. Date	2025-04

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.521	>1.510
S2	0.5	2.077	
S3	1.0	1.555	>0.930
S4	2.5	0.837	
S5	5.0	0.757	>0.450
S6	7.5	0.296	

✓ Approved / Not Approved  
✓ Quality Control

  
Signature

Date / 06-11-2023