

J.MITRA & CO. PVT. LTD.

TSH QUANTI MICROLISA

QC DATA SHEET

Kit	TSH QUANTI MICROLISA
Lot No.	ETH051023
Mfg. Date	2023-10
Expiry Date	2025-09

Standard	Concentration (μ IU/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	0.015	
S2	0.5	0.075	
S3	2.5	0.250	>0.150
S4	5.0	0.510	
S5	10.0	1.064	>0.640
S6	20.0	1.778	
S7	40.0	2.586	>1.550

Approved / Not Approved

Quality Control


Signature

Date 31-10-2023