

J.MITRA & CO. PVT. LTD.


T4 QUANTI MICROLISA

QC DATA SHEET

Kit	T4 QUANTI MICROLISA		
Lot No.	ETB010224		
Mfg. Date	2024-02		
Expiry. Date	2025-07		
Standard	Concentration ($\mu\text{g/dl}$)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.858	>1.710
S2	2.0	1.943	>1.170
S3	5.0	1.358	
S4	10.0	0.762	>0.460
S5	15.0	0.453	
S6	25.0	0.356	

Approved / Not Approved

Quality Control


Signature

Date: 02-03-2024