

J.MITRA & CO. PVT. LTD.

VITAMIN B12 QUANTI MICROLISA


QC DATA SHEET

Kit	VITAMIN B12 QUANTI MICROLISA
Lot No.	EVB010324
Mfg. Date	2024-03
Expiry. Date	2025-08

Standard	Concentration (ng/ml)	Avg. O.D. at 450/630 nm	Acceptable range of O.D.
S1	0	2.523	>1.510
S2	100	2.022	
S3	200	1.533	>0.920
S4	400	1.180	
S5	1000	0.629	>0.380
S6	2000	0.416	

✓
Approved / ~~Not Approved~~

Quality Control


Signature

Date 07-03-2024